



Statement of Conformity

with the internal metrology control programme of Fibre Lab and Co.

Internal metrology control programme - in-house method derived from IWTO-47

Statement no.	FL-2026-0001 (EXEMPLE)
Client	EXEMPLE - Élevage Fictif
Order / report reference	EXEMPLE CMD-0000 / RAPPORT-0000
Test date(s)	EXEMPLE 01/01/2026
Samples	EXEMPLE - 3 mèches d'alpaga (fictif)
Instrument	OFDA 2000, staple mode - commissioned in Saulzoir by the manufacturer (Mark Brims, OFDA)
Method	In-house method derived from IWTO-47
Reference document	Fibre Lab and Co. quality charter, version 3.0 (September 2026)

Fibre Lab and Co. states that the tests identified above were carried out in accordance with its internal metrology control programme, as described in its quality charter. The following checks were applied:

- CQ-1 - Built-in OFDA 2000 self-test on reference wires, at the start of every testing day
- CQ-2 - Check of the measuring bench: staple holder, lighting, movement
- CQ-3 - Reference staple measured for every batch, maximum accepted deviation of 0.3 μm on mean diameter
- CQ-4 - Run-to-run stability: 5 archive samples re-measured at least once per working week, maximum deviation of 0.3 μm
- CQ-5 - Samples conditioned to laboratory atmosphere before testing; mode and correction factor stated on the report
- CQ-6 - Traceability: unique identifier for each test, raw OFDA files retained
- CQ-7 - Re-measurement of archive samples: a set of archive samples re-measured at least twice a year and compared with our records
- CQ-8 - Stop rule: client testing suspended as soon as a CQ-1 to CQ-4 check is out of limits, resumed only after a compliant check

Scope

This statement covers only the samples received, as received, and the tests listed above. The results are given in the referenced test report, which alone is authoritative. It is not an assessment of the animal, the herd or any batch not tested.

Statement issued free of charge on request, within one week.

Responsible person	Bertrand Busin, Laboratory Manager
Issued at	Saulzoir
Date of issue	EXEMPLE 02/01/2026

Signature

